



Clifton Health Department Food Recall Notification



Food/Drug/Pet Recall: Novo Nordisk

Novo Nordisk is voluntarily recalling 1,468 product samples listed in the table below of Levemir®, Tresiba®, Fiasp®, Novolog® and Xultophy®, to the consumer level. These products are being recalled because they were stored at temperatures below storage requirements. This recall only impacts product samples and does not impact product that has been broadly distributed to pharmacies or mail-order services.

If product samples are exposed to temperatures below 32°F, it could cause a lack of efficacy and damage to the cartridge and pen-injectors. If product from an improperly stored vial, cartridge or pen-injector is used, there is a risk that you might not receive the right amount of medicine as intended which may lead to hyperglycemia or hypoglycemia resulting in adverse health consequences ranging from limited to life-threatening. Novo Nordisk has not received any reports of serious adverse events or injuries related to this recall.

These products are used to lower blood glucose levels in people with diabetes and are packaged in cartons with either a vial, pen-injector (FlexPen® or FlexTouch®) or a cartridge (PenFill®). A list of the affected lots can be found in the chart below:

Product Name	NDC #	Batch #	# of Affected Samples	Expiration Date
Fiasp® FlexTouch®	0169-3204-90 (Pen)	KP51207 24	06/30/2022	
	0169-3204-97 (Kit)	KP52618 153	10/31/2022	
Fiasp® PenFill®	0169-3205-91	KS6BF84 7	06/30/2022	
	KS6BX63 90		10/31/2022	
Fiasp® Vial	0169-3201-90	KS6AK76 10	05/31/2022	
	KS6BR92 20		09/30/2022	
	0169-6438-90 (Pen)	24		
Levemir® FlexTouch®	0169-6438-98 (Kit)	KP51933	07/31/2022	
	0169-6339-90 (Pen)	44		
NovoLog® FlexPen®	0169-5339-98 (Kit)	KS6BS11	11/30/2021	
NovoLog® Vial	0169-7501-90	JZFC826 17	06/30/2021	
	KZFM305 26		08/31/2022	
	JP52771 13		09/30/2021	
	JP53136 4		06/30/2021	
	KP50575 30		01/31/2021	
	KP50976 27		01/31/2022	
	KP51813 99		04/30/2022	
Tresiba® U100	0169-2660-90 (Pen)	KP52035 12	04/30/2022	
FlexTouch®	0169-2660-97 (Kit)	KP52117 36	04/30/2022	
	KP52440 207		06/30/2022	



Clifton Health Department Food Recall Notification



KP52461 60 04/30/2022
 KP52616 81 06/30/2022
 JP52361 7 08/1/2021
 KP52829 170 07/31/2022
 JP54181 12 09/30/2021
 Tresiba® U200 0169-2550-90 (Pen) KP51059 8 11/30/2021
 FlexTouch® 0169-2550-97 (Kit) KP51865 182 11/30/2021
 KP54179 68 11/30/2022
 JP52179 20 08/16/2021
 Tresiba® Vial 0169-2662-90 JZFE233 14 11/30/2021
 0169-2911-90 (Pen)
 Xultophy® Pen 0169-2911-97 (Kit) JP54291 3 06/20/2021

The product can be identified by looking for the batch number or lot number located on the product or carton and matching it to the list above. Novo Nordisk has notified all physician offices that received affected samples and requested all impacted samples be returned. Customers who received an affected sample through the physician's office should have received a letter from their physician. If product samples match a batch number above or there are any questions about the recall, please contact the Novo Nordisk recall processor Inmar at 1-888-686-5002, Monday through Friday, 9:00 AM to 5:00 PM EDT.

Please report any complaints and adverse events to Novo Nordisk's Customer Care Center which can be reached at 1-800-727-6500, Monday through Friday, 8:30 AM to 6:00 PM EDT.

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's MedWatch Adverse Event Reporting program either online, by regular mail or by fax.

- . Complete and submit the report Online
- . Regular Mail or Fax: Download form or call 1- 800-332-1088 to request a reporting form, then complete and return to the address on the pre-addressed form, or submit by fax to 1-800-FDA-0178

This recall is being conducted with the knowledge of the U.S. Food and Drug Administration.

No action is required of local health departments at this time for any of these recalls and alerts. If any requests for assistance are received from FDA, the Public Health and Food Protection Program will contact you. For additional information regarding warnings and recalls, please click on the weblink below.

For all recalls - <http://www.recalls.gov/recent.html>